

2nd Alternative test									
Laboratory ID	Treat used	Kid batch number	Expiry date	Geil Type (Phenex)	Batch	Expiry date	Positive control results		
418	10555 LIGHT PROTOCOL	01104	23/02/2022				3.268	3.267	0.042
528	DATA FROM BSE ELA Test Kit	021451	21/02/2022				3.825	3.824	0.014
							Negative control results		
							0.033	0.014	0.018
							0.013	0.014	0.014
							Cutoff		
							0.159	0.159	0.2
							Methods details		
							Maximum DO registered is 3.5. All data above are considered as "over"		

2nd Alternative test									
Intended Laboratory ID	Sample No. RB2101			Sample No. RB2102			Sample No. RB2103		
	NEGATIVE Bovine Tissue Homogenate			NEGATIVE Bovine Tissue Homogenate			C-POSITIVE Bovine Tissue Homogenate		
418	RUN			RUN			RUN		
	1st	2nd	3rd	1st	2nd	3rd	1st	2nd	3rd
528	0.126	0.021	0.023	0.067	0.028	0.002	3.207	3.460	3.460
	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve+++	Ve+++	Ve+++
Sample No. RB2104									
NEGATIVE Bovine Tissue Homogenate									
RUN									
1st	2nd	3rd	1st	2nd	3rd	1st	2nd	3rd	4th
0.054	0.031	0.019	0.054	0.021	0.021	0.054	0.021	0.021	0.017
Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-
Sample No. RB2105									
NEGATIVE Bovine Tissue Homogenate									
RUN									
1st	2nd	3rd	1st	2nd	3rd	1st	2nd	3rd	4th
0.054	0.031	0.019	0.054	0.021	0.021	0.054	0.021	0.021	0.017
Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-
Sample No. RB2106									
NEGATIVE Bovine Tissue Homogenate									
RUN									
1st	2nd	3rd	1st	2nd	3rd	1st	2nd	3rd	4th
0.054	0.031	0.019	0.054	0.021	0.021	0.054	0.021	0.021	0.017
Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-
Sample No. RB2107									
NEGATIVE Bovine Tissue Homogenate									
RUN									
1st	2nd	3rd	1st	2nd	3rd	1st	2nd	3rd	4th
0.054	0.031	0.019	0.054	0.021	0.021	0.054	0.021	0.021	0.017
Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-	Ve-

Sample No. RB2101 = Sample No. RB2102 = Sample No. RB2104 = Sample No. RB2105 = Sample No. RB2106 = Sample No. RB2107

Comments provided by Elena Bozzetta, IZS PLVA - Turin

Comments
no comments

CONCLUSION

All labs but one have passed the PT round. Well done.
Lab 655 failed the first round of the PT but passed the second one

The information contained in this Report must be considered confidential

Date:

Dr Elena Bozzetta,
TSE EUR/PTs Coordinator

END OF REPORT

Copies controlled