

Results Tabulation for Confirmatory Immunoblot test

MOD. N°	40IZ477	ED/REV. 00/01	STRUTTURA/UO:
PGS. N°	10DG010	PAGINA 1 di 3	DATA:



TSE EURL

Istituto Zooprofilattico Sperimentale del
Piemonte, Liguria e Valle d'Aosta – Turin

[Istituto Superiore di Sanità - Rome](#)

PROFICIENCY TESTING FOR VETERINARY LABORATORIES

Results Tabulation for Confirmatory Immunoblot test- Scheme: WB24	Report No. 13/2025	Distribution date: 10 th February 2025
---	--------------------	---

Table 1. General information about 'routine' methods applied by the participants in the WB24 PT

NRL ID code	Test Method used	Kit Batch Number	Kit Expiry Date	Gel Type	Gel Batch Number	Gel Exp. date	Positive Control 1 result	Positive Control 2 result	Negative Control 1 result	Negative Control 2 result	Negative Control 3 result	Negative Control 4 result	Positive Cut Off value
633	TeSeE Western Blot	4J0055	29/10/2025	CriterionTM XT Precast Gel	64617220	21/05/2025	2.670	2.734	0.013	0.014	0.014	0.013	0.201

Table 4. Results submitted by the participants in the WB24 PT applying 'routine' methods

NRL ID code	Test Method used	WB2406 Interpretation	WB2406 Characterization	WB2407 Interpretation	WB2407 Characterization	WB2408 Interpretation	WB2408 Characterization	WB2409 Interpretation	WB2409 Characterization	WB2410 Interpretation	WB2410 Characterization
Intended --->		Negative bovine tissue homogenate		C-positive bovine tissue homogenate		C-positive bovine tissue homogenate		Negative bovine tissue homogenate		Negative bovine tissue homogenate	
633	TeSeE Western Blot	negative	-	negative	-	Medium positive	-	negative	-	negative	-

Results Tabulation for Confirmatory Immunoblot test

MOD. N°	40IZ477	ED/REV. 00/01	STRUTTURA/UO:
PGS. N°	10DG010	PAGINA 2 di 3	DATA:



TSE EURL

Istituto Zooprofilattico Sperimentale del
Piemonte, Liguria e Valle d'Aosta – Turin

[Istituto Superiore di Sanità](#) - Rome

PROFICIENCY TESTING FOR VETERINARY LABORATORIES

Results Tabulation for Confirmatory Immunoblot test- Scheme: WB24	Report No. 13/2025	Distribution date: 10 th February 2025
---	--------------------	---

Comments provided by Maria Mazza, IZSPLVA - Turin (maria.mazza@izsplv.it)

WB2406 - Negative bovine tissue homogenate

The laboratory reported the correct result.

WB2407 - C-positive bovine tissue homogenate

The laboratory reported an incorrect result referring it as negative.

WB2408 - C-positive bovine tissue homogenate

The laboratory reported the correct result.

WB2409 - Negative bovine tissue homogenate

The laboratory reported the correct result.

WB2410 - Negative bovine tissue homogenate

The laboratory reported the correct result.

Results Tabulation for Confirmatory Immunoblot test

MOD. N°	40IZ477	ED/REV. 00/01	STRUTTURA/UO:
PGS. N°	10DG010	PAGINA 3 di 3	DATA:



TSE EURL

Istituto Zooprofilattico Sperimentale del
Piemonte, Liguria e Valle d'Aosta – Turin

[Istituto Superiore di Sanità](#) - Rome

PROFICIENCY TESTING FOR VETERINARY LABORATORIES

Results Tabulation for Confirmatory Immunoblot test- Scheme: WB24	Report No. 13/2025	Distribution date: 10 th February 2025
---	--------------------	---

Conclusions provided by Maria Mazza, IZSPLVA - Turin (maria.mazza@izsplv.it)

The laboratory also failed this second PT round and it has been contacted by the EURL TSE to discuss the situation and address any possible issues.

The information contained in this Report must be considered confidential

Date: 24th March 2025

Dr. Elena Bozzetta, Coordinator of PTs EURL TSE

Dr. Giuseppe Ru, Director of EURL TSE

Elena Bozzetta

Firmato digitalmente da:

Ru Giuseppe

Firmato il 03/04/2025 14:17

Seriale Certificato: 3441766

Valido dal 02/04/2024 al 02/04/2027

InfoCamere Qualified Electronic Signature CA

END OF REPORT

Copia controllata